

File No. 28(1119)/2016/Pur/T-264

Date: 06/03/2017

**Clarifications issued after the Pre-Bid Conference for the  
Procurement of Vibrational Circular Dichroism**

**CLARIFICATIONS**

**Firm Name: M/s. Anatek Services Pvt. Ltd.**

<b>Question</b>	<b>You have requested for the Fixed Liquid Cell with 100 <math>\mu\text{m}</math> path length but our principal supply only 150 <math>\mu\text{m}</math></b>
<b>Answer</b>	<b>Change of Fixed path length cell category 100 or 150 <math>\mu\text{m}</math> cell path (<math>\text{CaF}_2</math> Windows) instead of 100<math>\mu\text{m}</math> cell path (<math>\text{CaF}_2</math> Windows)</b>

**Tender Due Date & Time: 15<sup>th</sup> March 2017 up to 12.00 P.M (IST)**

**Technical Bid Opening Date & Time: 15<sup>th</sup> March 2017 up to 03.00 P.M (IST)**

**The above clarifications are part and parcel of the Tender Document.**

**Sd/-**

**STORES & PURCHASE OFFICER**